

Site d'Evry-Corbeil
Rue Henri Auguste Desbrières - BP 81
91003 Evry Cedex - France

O/Ref. : YQL/Basile 03605/14 LM

Phone : (33) 01.69.87.98.10

Evry, the April 3, 2014

Your contact : Mrs Marie-Noëlle HINARD
Mail : marie-noelle.hinard@snecma.fr

ELECTRON BEAM PROCESSES

Unit 4 Octimum

WOKING GU21 5SF

ROYAUME UNI

For the attention of Mr Raf ENGLEY
Chief Inspector

Subject : Laboratory approval (FAL n° 644)

Dear Mr,

After review of your quality organisation, technical capabilities and the survey of your Laboratory conducted by Mrs Marie-Noëlle HINARD the 19 of November 2013, we are pleased to confirm you the qualification of your laboratory for the test defined on the Special Processes Qualification Certificate (FAL n° 644) enclosed with the audit report and the Corrective Action Request n° CARs n° 200007353 (1) and 200007354 (2) closed.

Continuance of the laboratory qualification depends on conditions outlined in the Snecma procedures GRP-0087 and GRM-0092.

We would like thank you for the organization of this visit and for your cooperation.


Receive best regards.

The Auditor



Marie-Noëlle HINARD

The Quality Assurance Manager



Michel BIDEAU

Enclosed.: FAL 644
Audit report YQL n° 03605/14
CAR n° 200007353 (1) and 200007354 (2) closed
Nadcap Certificate
PXV 25521

Attestation de Qualification des Procédés Spéciaux

Special Processes Qualification Certificate

N° AQPS :
MDM503377
CODE FOURNISSEUR 023871-00
SUPPLIER NUMBER
FAL 644

SNECMA

Prononce la qualification du fournisseur suivant GRP-0087-GRM0092
Approval on the supplier qualification according to GRP-0087- GRM0092



ELECTRON BEAM PROCESSES


Unit 4 Octimum

WOKING GU21 5SF


ROYAUME UNI


Pour les procédés spéciaux suivant,
For the following special processes


Safran Process N°	Procédés Processes	Référentiel Technique Technical Specification	Statut	Restrictions techniques Technical limitation	Fin de Validité Expiration Date
L1e	Micrographie optique autres/ Other metallographic optical		Qualification	Micrographie des cordons de soudure FE suivant DMP 43002 et DMF Pièces	31/12/2017
N° de rapport Report n° BASE @ DOC BASILE N° YQL 03605/14 Auditeur/Responsable de la Qualification Auditor/Qualification Leader April 3, 2014					
Observations Remarks : Qualification provisoire prononcée sur l'audit Snecma/Qualification pronounced on Snecma audit					
		Marie-Noëlle HINARD - Michel BIDEAU			
		 			

		DEMANDE D'ACTION(S) CORRECTIVE(S) <i>(Corrective Action Request C.A.R)</i>		DAC N° 1 20000 7353 <i>[Identification CAR nr]</i>	
Intervention effectuée le : <i>(Intervention date)</i>		19/11/2013		Référence courrier: 46Ln° 03605/14	
Origine :		Initial audit		<i>[Ref.letter]</i> Date : 19/11/2013 <i>(Date issued)</i>	
Sujet : <i>(subject)</i>		Qualification Laboratory			
Nom et site du Fournisseur: <i>(Supplier's name and site)</i>		Electron Beam Processes ltd		N° Fournisseur: 02387100 <i>(Supplier Purchase code number)</i>	

Description de la non-conformité (Non-conformance description) <i>Thème ISO [ISO Them] 17025 §5.6</i> The magnifications of the two microscopes (Zeiss Axio and Leica M60) used for the metallographic welding inspection are not calibrated	
Classement : Ma C :critique(<i>critical</i>) Ma : majeure(<i>major</i>) mi : mineure(<i>minor</i>)	
Nom émetteur : YQL Marie-Noëlle HINARD A répondre pour le : 06/01/2014 <i>(Assessor)</i> <i>(Due date)</i>	

CAUSES racines (Root causes)		Méthodologie (Method)	
Although the microscopes were calibrated when the systems were set up and at each occasion when the software was up graded, they were not in the recall calibration system checks carried out against the certified master graticule was accepted by all other audit authorities and was deemed to be satisfactory. 1		Inspection 1.1	
ACTIONS CORRECTIVES (Corrective actions)		Application (Implementation) Date réalisation prévue <i>(planned completion)</i>	
Microscopes identified with S/N's EBP1004 & EBP1005. Microscopes now in the calibration system recall frequency set at 6Months calibration certs for microscopes and master graticule enclosed.		12/12/2013 3	
Nom du représentant de la société: Raf Engley <i>(Organization representative name)</i> 4		Signature :  Date : 12/12/2013 <i>(Current Date)</i> 6	

Vérification de la réalisation des actions par le Fournisseur <i>(Verification of the implementation of the completed corrective actions by the Supplier)</i>			
Nom du représentant de la société: Raf Engley <i>(Organization representative name)</i> 7		Signature :  Date : 20/01/2014 <i>(Current Date)</i> 9	

Vérification des actions correctives par l'émetteur / Commentaires : <i>(verification of the completed corrective actions by the assessor / Comment)</i>			
Date : <i>(Verification Date)</i>		Soldée Date : April 4, 2014 Nom : HINARD M.N. Clôturée Date : April 6, 2014 <i>(Name):</i>	
		Signature : <i>(Signature)</i> 	

CERTIFICATE OF CALIBRATION

Issued by

DERA CALIBRATION LABORATORIES

Part of DERA Defence Metrology Services

Date of Issue: 15 March 2001

Certificate Number:

200102525



0013

DERA

Defence Evaluation and Research Agency
DERA Calibration Laboratories
Johnson House
Risley
Warrington WA3 6AT

Tel: +44 (0) 1925 253229
Fax: +44 (0) 1925 252343

Page 1 of 2

APPROVED SIGNATORY

Chris Miller/Brian Bowden

Customer : Pyser SGI Limited
Address : Kent, TN8 6HA
Order Number : 66886
Date Received : 13 March 2001
Date Calibrated : 14 March 2001
Manufacturer : Graticules
Type : S78
Serial Number : SC 610
Description : Stage Micrometer Glass Scale
Calibrated By : B Bowden

REPORT

- 1 The Glass Scale was calibrated on the MUL 214B Measuring Machine with Hewlett Packard Laser Interferometer attached (Cert No. 08A 011/9117 NPL).
- 2 The results are tabulated on page 2.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to recognised standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standard laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

DERA is an Agency of the UK Ministry of Defence

E.B.P. CALIBRATION REPORT

CERTIFICATE No EBP CAL 1000

MICROSCOPE S/N EBP1004
MASTER GRATICLE S/N SC610
CALIBRATED AGAINST 1000 μ m REFERENCE DIMENTION (M.P.E. 5%)




Image No. Calibration - 20110302_000018
Comment:
CALIBRATION IMAGE 1.0 - 12.12.2013




Image No. Calibration - 20110302_000019
Comment:
CALIBRATION IMAGE 1.6 - 12.12.2013




Image No. Calibration - 20110302_000020
Comment:
CALIBRATION IMAGE 2.5 - 12.12.2013




Image No. Calibration - 20110302_000021
Comment:
CALIBRATION IMAGE 4.0 - 12.12.2013

CALIBRATION DATE : 12/12/2013
DUE DATE :12/06/2014
CALIBRATED By Raf Engley

E.B.P. CALIBRATION REPORT

Page

1/1

CERTIFICATE No EBP CAL 1001

MICROSCOPE S/N EBP1005

MASTER GRATICL S/N SC610

CALIBRATED AGAINST 1000 μ m REFERENCE DIMENTION (M.P.E. 5%)




Image No. Calibration - 20110302_000023

Comment:

CALIBRATION IMAGE 5X - 12/12/2013 = 1000 μ m




Image No. Calibration - 20110302_000024

Comment:

CALIBRATION IMAGE 10X - 12/12/2013 = 1000 μ m




Image No. Calibration - 20110302_000025

Comment:

CALIBRATION IMAGE 20X - 12/12/2013 = 500 μ m




Image No. Calibration - 20110302_000026

Comment:

CALIBRATION IMAGE 50X - 12/12/2013 = 200 μ m

CALIBRATION DATE : 12/12/2013


DUE DATE : 12/06/2014

CALIBRATED By Raf Engley

Title: Calibration of Microscopes

Electron Beam Processes Limited



Equipment


Inspection master graticule Serial No.SC610

Method

1. Using Imagic software capture and image of the graticule scale and measure it over its 1000 μ m length or as long as the magnification allows.
2. Measurements to be taken at four set magnification levels to complete the calibration.
3. Create calibration report containing the four images and save in swap\Inspection\CALIBRATION\Calibration Records-Tools-Labels\EBP CAL CERTS.
4. Calibration frequency to no longer than six months.

Approved:

File Name	Page	Revision	Date
PPI 8608	1 OF 1	1.0	12/12/13

		DEMANDE D'ACTION(S) CORRECTIVE(S) <i>(Corrective Action Request C.A.R)</i>		DAC N° 2 200007354 <i>[Identification CAR nr]</i>	
Intervention effectuée le : <i>(Intervention date)</i>		19/11/2013		Référence courrier: 46Ln°03605/14	
Origine :		Initial audit		<i>[Ref.letter]</i> Date : 19/11/2013 <i>(Date issued)</i>	
Sujet : <i>(subject)</i>		Qualification Laboratory			
Nom et site du Fournisseur: <i>(Supplier's name and site)</i>		Electron Beam Processes Ltd		N° Fournisseur: 023871-00 <i>(Supplier Purchase code number)</i>	

Description de la non-conformité (Non-conformance description)

Thème ISO [ISO Them] 17025 §5.6

For the inspection of the welding samples, EBP uses a form document (E/5535-5 : Batch control- linear test piece-case,HP turbine). In this document :

- One criteria of the DMP43-002 is missing (Lmean-0,15 <L< Lmean+0,15)
- The is one mistake (DMD43-002 instead of DMP43-002)
- The documents for the preparation and inspection of the welding ("Inspection procedure for Snecma batch control parts" and "Production process- Silvercrest HP turbine case linear test piece" are not mentionned
- These both documents have no reference number and one of them "Production process- Silvercrest HP turbine case linear test piece" has no date and no issue

Classement : mi C :critique(critical) Ma : majeure(major) mi : mineure(minor)

Nom émetteur : YQL Marie-Noëlle HINARD A répondre pour le : 06/01/2014
(Assessor) (Due date)

CAUSES racines (Root causes)

Méthodologie (Method)

Documents that were part of the job pack and therefore stored together were not deemed requireing of seperate indentification as they were not use with anything else.

1

planning

1.1

ACTIONS CORRECTIVES (Corrective actions)

Application (Implementation)

Date réalisation prévue
(planned completion)

(Lmean-0.15<L<Lmean+0.15) included in report format
Inspection procedure now identified as IP2010 on routing
Production procedure-Silvercrest now identified on routing as SNECMA-PP-5535BLTP with date and issue.
Copies of routings, procedures and report sample enclosed.

2

20/01/2014

3

Nom du représentant de la société: Raf Engley
(Organization representative name)

4

Signature :



Date : 20/01/2014
(Current Date)

6

Vérification de la réalisation des actions par le Fournisseur

(Verification of the implementation of the completed corrective actions by the Supplier)

Nom du représentant de la société: Raf Engley
(Organization representative name)

7

Signature :



Date : 20/01/2014
(Current Date)

9

Vérification des actions correctives par l'émetteur / Commentaires :
(verification of the completed corrective actions by the assessor / Comment)

Date

Soldée

Date : April 4, 2014

Nom : HINARD M.N

Signature :


(Verification Date)

Clôturée

Date : April 4, 2014

(Name):

(Signature)



COMPTE RENDU D'AUDIT

Nom du Fournisseur <i>Supplier</i>	Electron Beam Processes	Code fournisseur <i>Supplier number</i>	023871-00		
Adresse <i>Address</i>	Unit 4 Octimum - WOKING GU21 5SF - United Kingdom				
Activité Fournisseur <i>Supplier Activities</i>	Beam welding supplier				
Contact fournisseur <i>Supplier contact</i>	Nom R. ENGLE		Function Chief Inspector		
	Adresse email Email adressa : r.engley@ebpglobal.com				
Nature Audit <i>Audit Purpose</i>	Sur site <i>On-site audit</i>	<input checked="" type="checkbox"/>	Sur Dossier(DQ) <i>On DQ</i>	<input type="checkbox"/>	Date : November 19th, 2013
	Evaluation <i>Evaluation</i>	<input checked="" type="checkbox"/>	Renouvellement <i>Re-Approval</i>	<input type="checkbox"/>	Qualification initiale <i>Approval</i>
					<input checked="" type="checkbox"/>
	Surveillance <i>Follow up</i> <input type="checkbox"/> Autre* <i>Other</i> <input type="checkbox"/>				
*Si Autre préciser : <i>If other detail :</i>					
Thème Audit <i>Scope of Audit</i>	Lab audit				

Equipe audit & audités <i>Audit staff & auditees</i>				
Fonction <i>Title</i>	Nom <i>Name</i>	Safran	Fournisseur <i>Supplier</i>	Catégorie de l'auditeur Safran <i>Safran auditor level</i>
Responsable of the audit	Marie-Noelle HINARD			Safran auditor
Auditee	Raf ENGLE		Chief inspector	
Auditee	Kenneth HOWLETT		Planning and quality Engineer	
Auditee	Jeff BOYES		Operations Executive	

Spécification(s) de l'audit <i>Audit Specification</i>	
Référentiels techniques Safran <i>Safran Technical Specification</i>	DMP 43-002
Autres référentiels <i>Other specification</i>	ISO 17025

COMPTE RENDU D'AUDIT

Points Forts <i>Positive points</i>	Points Faibles <i>Weak points</i>
<ul style="list-style-type: none"> - EBP has a NADCAP approval. For the welding and NDT - Technical level of the tests is good - Every operator is qualified - Quality system (Management review - planning of audits management of the documents) 	<ul style="list-style-type: none"> - Calibration of the magnifications of the microscopes (CAR n° 200007353 (1) Major) - Documents on the batch control linear test piece (Car n° 200007354 (2) – minor) - Addition of one line on the form of the job order for the inspection of the batch sample <p>Other points :</p> <p><u>Point 1 :</u></p> <p>Before to give the qualification, Snecma ask to Smith the five last reports of the batch control linear test piece and the sample to perform the checking in his lab. Snecma will give the qualification :</p> <ul style="list-style-type: none"> - After reception of the answers of the two findings - If there is a correlation between Snecma results and EBP results <p><u>Point 2 :</u></p> <ul style="list-style-type: none"> - Snecma will send to EBP :the DMC0090 for the metallographic inspection with the list of the reagents - The number of ASTM standard for the calibration of the microscope

Eléments analysés <i>Items reviewed</i>			
1	AQM-13 - Annual Quality Management review minutes – date 15/01/2013	6	Document : Inspection procedure for Snecma batch control part
2	EBP test report n°A01360_000204	7	
3	Quality manual	8	
4	Form document : E/5535-5	9	
5	Document : Production process- Silvercrest HP turbine case linear test piece	10	

COMPTE RENDU D'AUDIT

Tableau des écarts Deviation table						
N° de l'écart Deviation n° JJ	Niveau Level	Procédé Process	Description de l'écart Non Conformance Report Responsible	Responsable Responsible	Délai Time limit	Statut Status
DAC 200007353 (1)	M	Calibration of the microscope	<i>The magnifications of the two microscopes (Zeiss Axio and Leica M60) used for the metallographic welding inspection are not calibrated</i>	Chief inspector	06 th January 2014	
DAC 200007354 (2)	m	Checking of the welding following DMP 43-002	For the inspection of the welding samples, EBP uses a form document (E/5535-5 : Batch control- linear test piece-case,HP turbine). In this document : - One criteria of the DMP43-002 is missing (Lmean-0,15 <L< Lmean+0,15) - The is one mistake (DMD43-002 instead of DMP43-002) - The documents for the preparation and inspection of the welding ("Inspection procedure for Snecma batch control parts"and "Production process- Silvercrest HP turbine case linear test piece" are not mentionned - These both documents have no reference number and one of them "Production process- Silvercrest HP turbine case linear test piece" has no date and no issue	Chief inspector	06 th January 2014	

m = mineur/ minor M = majeur/ major

Tableau des Recommandations Préventives table		
N° de la recommandation Preventive n°	Procédé Process	Description Description
1	Inspection of beam welding	Mention one line concerning the welding sample inspection on the work order to avoid any risk
2	Inspection of beam welding	Mention the reference number of the microscope on the report of the welding sample

Audit Notes**1) Context**

- Initial audit of qualification
- EBP has a NADCAP approval for electron beam welding

2) Results of the System Audit

- Le lab depends of the inspection department. The chief is Raf Engley
- Date of the last revision of the quality manual : QP7 – 06/12
- There is one management review per a year – date for the last one : 15/01/2013

Management of documents:

- A procedure exists – QM7
- A procedure "Process and procedure Instruction n° 8562" describes the processing of enquiries and creation of production Paperwork

Management of supplier's documents:

- One computerized list of the SAFRAN documents exists. It is updated when a new document arrives - Example : DMP43-002 has the good issue

Audit :

- A planning of audits exists. The report of the last audit has been seen during the audit

Last NADCAP audit: 22 April 2013

- 3 minor Cars but not for the lab

Management of the preventive actions, corrective actions and complains:

This point is mentioned by the quality manual – paragraph 8.5 and procedure QM 8.9 and Quality
One complain was seen during this audit.

There is no indicator on the non conformities but the management review mentions them

Subcontracting of tests and calibrations:

The calibration of the calipers that EBP used for the measurement of the welding samples is subcontracted every 6 months – Supplier Brunel

3) Skills and qualification :

Every operator and the chief were trained by Snecma

Survey of the implementation of the test

For the inspection of the welding samples, EBP uses a form document (E/5535-5 : Batch control- linear test piece-case,HP turbine). In this document :

- One criteria of the DMP43-002 is missing ($L_{mean}-0,15 < L < L_{mean}+0,15$)
- There is one mistake (DMD43-002 instead of DMP43-002)
- The documents for the preparation and inspection of the welding ("Inspection procedure for Snecma batch control parts" and "Production process- Silvercrest HP turbine case linear test piece" are not mentioned
- These both documents have no reference number and one of them "Production process- Silvercrest HP turbine case linear **test piece**" **has no date and no issue**

→ **DAC n° 200007354 (2)**

An instruction procedure exists for the Snecma batch control . In this document it is explained in detail specially how to prepare the longitudinal sample.

If the result is good, the lab gives the green light for the welding parts but **on the work order, this is no line for the inspection of the welding sample → improvement : include a line on the work order for the checking of the welding sample**

During the audit, one sample was prepared and examined → satisfactory

For the longitudinal sample, a caliper is used. It is calibrated every 6 months – its certificate was seen during the audit

Calibration of the microscopes: The lab has a micrometer with a certificate but the magnifications of the two microscopes (Zeiss Axio and Leica M60) used for the metallographic welding inspection are not calibrated → DAC n° 200007353 (1)

Archiving

After the checking of the welding samples, EBP sends them and the reports to Smith – Snecma supplier of the cases

Cross tests between EBP lab and Snecma Corbeil lab : see PVX 25521

These cross tests were been done on four samples – see the list after

- HPT case 14 and 15 : S/N DE952327-3 and DE952328-2 (report EBP n°A01360-000157)
- HPT case 16 and 17 : S/N DE952329-1 and DE952330-R (report EBP n°A01360-000190)
- HPT case 18 and 19 : S/N DE952331-P and DE952332-N (report EBP n°A01360-000204)
- HPT case 20, 23 and 24: S/N DE952333-M, DE952334-L and DE952335-K (report EBP n°A01360-000225)

The results obtained by SNECMA laboratory are similar to the results obtained by EBP laboratory. The geometrical inspection and the checking for internal anomalies are in conformity with the requirements of the DMP43-002 index E.

Note: Great care should be taken to the identification of the samples. In fact, pen-writing was sometimes different to the marking engraved on the back side of the mounting